

CERTIFICATE OF ANALYSIS



1	Name of product	Blood Bag CPDA-1; 150 ml; Single Bag;		
2	Description of product	SINGLE Blood bag with donor tube and donor needle of 16G; with Needle Injury Protector, 1 bag in an Aluminium pack.		
3	Batch No.	01F22Y1		
4	Mfg. Date	2022/06		
5	Expiry Date	2025/05		
6	Product Code	PB-1CD156M5S		
7	Batch Size	1860 Nos.		
8	Quantitative assay per dosage form			
	Ingredients (g/L)			
	CPDA-1 Solution – USP		Acceptance Limit	Assay Result
	a	Monobasic Sodium Phosphate (Monohydrate)	2.11 - 2.33	2.18
	b	Dextrose (Monohydrate)	30.30 - 33.50	31.47
	c	Total Citrate (as Citric acid anhydrous)	19.16 - 21.18	19.75
	d	Sodium	6.21 - 6.86	6.47
9	Other tests or requirements			
	CPDA-1 Solution – USP		Acceptance Limit	Result
	a	Chloride	Not more than 35 ppm	Passed
	b	pH	5.0 – 6.0	5.68
	c	Volume of Anticoagulant solution	20 – 23 mL	23
	d	Bacterial Endotoxin Test	Not more than 5.56 EU/mL	Passed
	e	Sterility	No growth of microorganisms	Passed
f	Particulate matter	Shall comply USP <788>	Passed	
10	Functional test for PVC container according to ISO 3826-1: 2019	Shall comply with ISO 3826-1	Passed	
Judgment : Complies with USP 43 & ISO 3826-1:2019				
Date : 17/06/2022 (DD/MM/YYYY)				
Prepared By: V.V.Kumari Deputy Manager (QC)		Approved By: Sreejith Nemathillam Analytical Chemist		

TERUMO PENPOL PRIVATE LIMITED, CIN: U33112KL1985PTC004531, E-Mail: Penpol.info@terumobct.com

Registered Office : I-2, Jawahar Nagar, Kowdiar P.O, Thiruvananthapuram - 695 003 Kerala, India. Phone: +91 471-3015500, +91 471-3015501
Blood Bag Factory : VP 1/520, Puliyaikonam P.O, Thiruvananthapuram - 695 573, Kerala, India. Phone: +91 471-3052022 / 7192210 /3015600 / 3015606 / 3015609
Medical Systems Group : T.C. 27/373, Andoor Buildings, General Hospital Road, Thiruvananthapuram - 695 035 Kerala, India. Phone: +91 471-7135800

FR-QCD-77/A

LOT CERTIFICATE

1	Name of product	Blood Bag CPDA-1; 150 ml; Single Bag;	
2	Description of product	SINGLE Blood bag with donor tube and donor needle of 16G; with Needle Injury Protector, 1 bag in an Aluminium pack.	
3	Lot No./Batch No.	01F22Y1	
4	Mfg. Date	2022/06	
5	Expiry Date	2025/05	
6	Product Code	PB-1CD156M5S	
7	Lot Size	1860 Nos.	
8	Intended use	Blood bag is a sterile medical device for single use only and used for collecting whole blood as well as for separation & storage of blood components.	
9	Obtained certificates	CE Mark-Approval of EC Directive 93/42/EEC on medical devices, Annex II (excluding section 4), Certificate no. IN19/818843705 issued by SGS Belgium NV, Notified Body 1639. Quality Management System BSI ISO 9001:2015 certificate number FM747384 issued by BSI Assurance UK Limited, England, UK & IN06/70701 (ISO 13485) issued by SGS United Kingdom Ltd, Under UKAS accreditation.	
10	Type of sterilization	Moist heat sterilization and the sterilization process has been validated and is controlled as specified in ISO-17665-1 with a SAL of 10 ⁻⁶ .	
11	Material validation	Quality of plastics is validated according to ISO 3826 and United States Pharmacopoeia.	
QC specification and inspection result			
12	a	Chemical test for anticoagulant solution as per USP	Passed
	b	Functional test for PVC container according to ISO 3826	Passed
	c	Biological test (Bacterial Endotoxin Test) according to USP	Passed
	d	Microbiological test (Sterility test) according to USP	Passed
Judgment : Complies with USP 43 & ISO 3826-1:2019			
Date : 17/06/2022 (DD/MM/YYYY)			
Prepared By: V.V.Kumari Deputy Manager (QC) 		Approved By: Sreejith Nemathillam Analytical Chemist 	

TERUMO PENPOL PRIVATE LIMITED, CIN: U33112KL1985PTC004531, E-Mail: Penpol.info@terumobct.com

Registered Office : I-2, Jawahar Nagar, Kowdiar P.O, Thiruvananthapuram - 695 003 Kerala, India. Phone: +91 471-3015500, +91 471-3015501
Blood Bag Factory : VP 1/520, Puliyaikonam P.O, Thiruvananthapuram - 695 573, Kerala, India. Phone: +91 471-3052022 / 7192210 /3015600 / 3015606 / 3015609
Medical Systems Group : T.C. 27/373, Andoor Buildings, General Hospital Road, Thiruvananthapuram - 695 035 Kerala, India. Phone: +91 471-7135800

TERUMOPENPOL.COM