

DECLARATION OF CONFORMITY

No. ADAMS_005-D-23136

Manufacturer: ARKRAY Factory, Inc.
Address: 1480 Koji, Konan-cho, Koka-shi, Shiga 520-3306, JAPAN
Single Registration Number: JP-MF-000003008

Authorised Representative: ARKRAY Europe, B.V.
Address: Prof. J.H. Bavincklaan 2, 1183 AT Amstelveen, THE NETHERLANDS
Single Registration Number: NL-AR-000002056

Name(s) of the Device(s), Product code(s)/Catalogue number(s), Classification(s), Basic UDI-DI(s), EMDN code(s), Serial/Lot number(s), and Intended purpose(s): Refer to the attachment

<IVD Regulation>

Notified Body: N/A

Conformity assessment route: The following procedures are used for the CE-labeling of our products according to the Regulation IVDR 2017/746:
Class A:
ANNEX II and ANNEX III are applied

Common Specifications: N/A

EU Technical Documentation Assessment Certificate: N/A
EU Quality Management System Certificate: N/A

This declaration of conformity is issued under the sole responsibility of ARKRAY Factory, Inc. We hereby declare that the medical device(s) specified in the attachment meet(s) the provisions of the Regulation (EU) IVDR 2017/746 for in vitro diagnostic medical devices. This declaration is supported by the Quality System approval to EN ISO 13485: 2016 issued by TÜV SÜD Product Service GmbH. All supporting documentation is retained at the premises of the manufacturer.

Shiga, JAPAN

2023-05-30

Place and Date

Shuzo Joko

Shuzo Joko
Person Responsible for Regulatory Compliance
ARKRAY Factory, Inc.

No. ADAMS_005-D-23136

Attachment

< Device(s) information >

#	Name	Product code / Catalogue number	Classification	Basic UDI-DI	EMDN code	Serial/Lot number
1	HEMOLYSIS WASHING SOLUTION 90H	107374	A	4987486BC0107374H3	W01019001	Next number from 3C3222

< Intended purpose(s) >

1. HEMOLYSIS WASHING SOLUTION 90H is a washing solution for use with HA-8190V, which is intended for the quantitative measurement of HbA1c. For *in vitro* diagnostic use and professional use only.