

Certificate

Traceability of Manufacturers to the IFCC Reference Measurement Procedure for HbA1c

This certifies that ARKRAY using ADAMS A1c HA-8180V, uses calibrators supplied by the IFCC Network to get traceable to the IFCC Reference Measurement Procedure and participates in the Monitoring Programme to demonstrate traceability. In the Monitoring Programme of 2013 the following performance was seen:

Deviation from IFCC-target	at 30 mmol HbA1c/mol Hb :	0.3
	at 60 mmol HbA1c/mol Hb :	0.3
	at 90 mmol HbA1c/mol Hb :	0.2
Reproducibility, coefficient of variation		0.50%
Linearity, correlation coefficient		0.9997

Date of issue: 5 December 2013

Certification expires: 31 December 2014



IFCC Network Coordinator