



检验报告单

CERTIFICATE OF ANALYSIS

SJJ BG/DR-0605-012-01

产品名称：尿液分析用鞘液

Product Name: Sheath II

生产批号：20221015

生产日期：20221015

失效日期：20240414

检验日期：20221015

Production Lot No.:

Date of Manufacture:

Expiry Date:

Test Date:

序号 No.	检验项目 Test Item	标准规定 Criteria	检验结果 Test Result
1	外观 Appearance	无色透明液体，不得有沉淀、颗粒、絮状物。 Colorless, transparent liquid; no sediments, particles or floccules	合格 Qualified
2	PH 值 PH	7.55±0.20 (25°C±1)	7.57
3	电导率 Conductivity	(13.80±0.50) mS/cm (25°C±1)	13.99
4	净含量 Net Volume	不少于标示值 No less than the labeled value	合格 Qualified
5	渗透浓度 Osmotic Concentration	(275±10) mmol/L (275±10mOsm/kg)	277
6	包装 Package	外包装和说明书的文字内容、尺寸应与文审审批稿一致， 标签、包装盒上批号及失效日期符合生产计划要求 The package and instructions contents and dimensions should be consistent with the approved document review sheet; the lot number and expiry date on the label and package should meet the requirements of the manufacture plan.	合格 Qualified
结论 Conclusion	符合标准，准许出厂 The criteria are met; the product is permitted to leave the factory.		

检验员：

齐盼盼

Tested by:

复核人：

郭小敏

Checked by:

批准放行人：

齐琳

Approved by: