

## Certificate of Analysis

Cat No.	Product Name	
BC1110Z	Seeplex® Leukemia BCR/ABL	
Lot#	BC0122D02	
Date of Manufacture	2022-04-08	
Expiry Date	2023-04-07	
Package Size	25 rxns	
Storage	Stored at -20°C	

components	Specification	Results(PASS/FAIL)
Check of nonspecific band	PCR rxn & Electrophoresis (or Auto capillary electrophoresis)	PASS
Marker Intensity Test	Electrophoresis (or Auto capillary electrophoresis)	PASS
Appearance	Check: Label Lot No. Cat No. Appearance of box Components Manual book	PASS
Check on contamination	PCR rxn & Electrophoresis (or Auto capillary electrophoresis)	PASS
Sensitivity Test	PCR rxn & Electrophoresis (or Auto capillary electrophoresis)	PASS
Positive Control Test	PCR rxn & Electrophoresis (or Auto capillary electrophoresis)	PASS

<sup>\*</sup> This Lot was inspected according to Seegene's Quality Control Instruction, and we confirm that all of the results meet the acceptance criteria. This CoA is vaild for 1 year from issued date.

(Issued date) 2022-12-13

(Signature)

Head of Quality Assurance Department

Seegene Inc. Taewon Bldg., 91. Ogeum-ro, Songpa-gu, Seoul,

05548, Korea TEL :(82) 2-2240-4000 FAX :(82) 2-2240-4040 email : cs@seegene.com

1page QF-0803-06('21.03.05) Rev.3