



Certificate of Analysis

| | |
|----------|--------------------------|
| Cat No. | Product Name |
| RV10307X | Allplex™ RV Master Assay |

| | |
|---------------------|-----------------|
| Lot# | RVF522H02 |
| Date of Manufacture | 2022-08-05 |
| Expiry Date | 2023-09-04 |
| Package Size | 100 rxns |
| Storage | Stored at -20°C |

| components | Specification | Results(PASS/FAIL) |
|---------------------------|--|--------------------|
| Check of nonspecific band | PCR rxn & Electrophoresis (or Auto capillary electrophoresis) | PASS |
| Check on contamination | Real-Time PCR rxn | PASS |
| Internal Control Test | Real-Time PCR rxn | PASS |
| Sensitivity Test | Real-Time PCR rxn | PASS |
| Negative Control Test | Real-Time PCR rxn | PASS |
| Positive Control Test | Real-Time PCR rxn | PASS |
| Appearance | Check: Label Lot No. Cat No. Appearance of box Components Manual book | PASS |

* This Lot was inspected according to Seegene's Quality Control Instruction, and we confirm that all of the results meet the acceptance criteria. This CoA is valid for 1 year from issued date.

(Issued date) 2023-02-24

(Signature)

Head of Quality Assurance Department

Seegene Inc. Taewon Bldg., 91. Ogeum-ro, Songpa-gu, Seoul,
05548, Korea
TEL : (82) 2-2240-4000
FAX : (82) 2-2240-4040
email : cs@seegene.com