

BELDICO SA

Rue André Feher, 3
Zoning de Aye
6900 Marche-en-Famenne
BELGIUM

Date:16/03/2018

CERTIFICATE OF CONFORMITY

We declare that the following products:

Code: HPF0612
Batch no: 18C09

Name: PRODUCT SAMPLING

are manufactured by Beldico, in Belgium, in conformity with our internal specifications, working procedures and our quality system ISO 9001-2008, ISO 13485-2003 and in accordance with the Directive for Medical Devices 93/42/EEC.

The above-mentioned items are Gamma-irradiated, in conformity with norm ISO11137

Sterilisation date: 15/03/18
Sterilisation number: I25181001

and have been declared sterile according to the conformity of the physical sterilisation parameters.

QA/Pharmacist



Tel.: +32 (0)84/31 37 80
Fax: +32 (0)84/31 37 87
E-mail: info@beldico.be
Website: <http://www.beldico.be>

RC/HR: Marche-en-Famenne 14917
TVA/BTW: BE 432 680 178
BBL 310-0669414-06

