

CERTIFICATE OF CONFORMITY INTERCEPT PRODUCT



Formulaire N°: C22-TEST-F

: B Révision N°

SOP N° : TEST-F

Page 1 of 1

CODE	DDE INT3104B		CE19G12L72
		Expiry Date	2021-06

		Lot	Raw material batch		
Amotosalen solution sub-	1	CE19H01L11	Amotosalen	1	180093
assembly	2	N/A		2	N/A
	3	N/A		3	N/A
PCT sub-assembly	1	CE19G12L72	CAD material		81112A1 / 81119A1 / 81119A3 /
					81126A1 / 81203A1

DESCRIPTION: INTERCEPT Processing Set for Plasma

CONFORMITY OF TESTS RESULTS

	PYROGENICITY TEST: Turbidimetric Kinetic (TEST-B-03)					
Product	Limits (≤)	Date Passed		Results		
Final product	0.125 EU / mL	August 7, 2019		PASS		
Amotosalen solution sub-assembly	□ 3 mM → 2.0 EU / mL	1	August 8, 2019	PASS		
	OR	2	N/A	N/A		
	☑ 6 mM → 0.8 EU / mL	3	N/A	N/A		
PCT sub-assembly	0.25 EU / mL	1	July 17, 2019	PASS		

Disposition : PASS

Disposition: PASS

Disposition: PASS

BACTERIOLOGY-BIOBURDEN (TEST-B-17 / TEST-B-09) Disposition: PASS FINAL PHYSICAL TESTING (TEST-C) Disposition: PASS pH ANALYSIS ON CONNECTED KIT (TEST-A) Disposition: PASS

AMOTOSALEN FINAL CHEMICAL ANALYSIS (TEST-A-01)

Steam sterilization cycle according to validated method reviewed & conform	CONFORM ☑
Gamma irradiation according to validated method reviewed & conform	CONFORM ☑
E-Beam irradiation according to validated method reviewed & conform	CONFORM ☑
No positives	CONFORM ☑

Note: N/A means Not Applicable PCT means Photo Chemical Treatment

QUANTITY

2592 UNITS

Reviewed by Fenwal QA MANAGER or DELEGATE Date & Signature

Released by Cerus QA MANAGER or DELEGATE Date & Signature

C. BERTHOMIER

STERILITY

Amotosalen solution

PCT sub-assembly

Connected product

Sterilility Test (TEST-B-13)

September 11, 2019

Romain LELIEVRE Sept 11th 2019

This is to certify that this product was manufactured according to current Good Manufacturing Practice and fulfills the requirements of the Device Master Record.