



**JMS SINGAPORE PTE LTD**  
440 Ang Mo Kio Industrial Park 1, Singapore 569620  
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)  
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)  
Web : www.jmss.com.sg  
Reg : 197901579C

Page 1 of 1

TO MEDISTA SPOL S.R.O.

**Certificate of Compliance**

**WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS**

**INVOICE NO** 1018348  
**DATE OF ISSUE** 29-DEC-17

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-2577	BLOOD BAG CPDA-1 (S) 250ML IBSP-NP-CLP 16G CE	171025003	2,800	25/10/17	OCT2020
811-8442	BLOOD BAG CPD-SAGM (Q) 450+150+400MLX2 5DPC IBSP-NP-BSB-CLP 16G (DT) CE	171019004	1,080	19/10/17	OCT2020
811-8475	BLOOD BAG CPD-SAGM (Q) 450+150+400MLX2 5DPC NP-CLP 16G CE	171215006	1,008	15/12/17	DEC2020
814-0135	TRANSFER BAG 150ML W/6 BAGS W/O SERIAL NUMBERS W/CLAMPS CE	1711108001	1,020	08/11/17	NOV2020
814-0302	TRANSFER BAG 300ML CE	171018002	1,000	18/10/17	OCT2020
814-0503	TRANSFER BAG 500ML CE	171127001	1,000	27/11/17	NOV2020
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	171207901	2,016	07/12/17	DEC2020

TEST ITEMS	JUDGEMENT	REMARKS
1. TOTAL APPEARANCE	PASSED	
2. FUNCTION TEST	PASSED	
3. PYROGENITY-USP	PASSED	
4. STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS  
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

*fb*  
CY CHIA  
QA MANAGER

P.O. NO : 2170376,2170381