

**JMS SINGAPORE PTE LTD**

440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

Page 1 of 1

TO MEDISTA SPOL S.R.O.

Certificate of Compliance**WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS****INVOICE NO** 1019441 **DATE OF ISSUE** 20-NOV-18

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-8442	BLOOD BAG CPD-SAGM (Q) 450+150+400ML X2 5DPC IBSP-NP-BSB-CLP 16G (DT) CE	181017002	1,836	17/10/18	OCT2021
811-8442	BLOOD BAG CPD-SAGM (Q) 450+150+400ML X2 5DPC IBSP-NP-BSB-CLP 16G (DT) CE	181018002	1,764	18/10/18	OCT2021
814-0302	TRANSFER BAG 300ML CE	181012001	1,000	12/10/18	OCT2021
831-8307	BLOOD BAG T&B CPD-SAGM (T) 450+400ML X2 IBSP-NP-BSB-CLP 16G (DT) CE	181020001	4,000	20/10/18	OCT2021

TEST ITEMS	JUDGEMENT	REMARKS
1. TOTAL APPEARANCE	PASSED	
2. FUNCTION TEST	PASSED	
3. PYROGENITY-USP	PASSED	
4. STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

P.O. NO : 2180356

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

CY CHIA
QA MANAGER