



**JMS SINGAPORE PTE LTD**  
440 Ang Mo Kio Industrial Park 1, Singapore 569620  
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)  
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)  
Web : www.jmss.com.sg  
Reg : 197901579C

Page 1 of 1

TO MEDISTA SPOL S.R.O.

**Certificate of Compliance**

**WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS**

**INVOICE NO** 1020691 **DATE OF ISSUE** 26-NOV-19

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-8442	BLOOD BAG CPD-SAGM (Q) 450+150+400MLX2 5DPC IBSP-NP-BSB-CLP 16G (DT) CE	191107001	3,096	07/11/19	NOV2022
811-8442	BLOOD BAG CPD-SAGM (Q) 450+150+400MLX2 5DPC IBSP-NP-BSB-CLP 16G (DT) CE	191108005	2,304	08/11/19	NOV2022
814-0503	TRANSFER BAG 500ML CE	191009001	1,000	09/10/19	OCT2022
814-0504	TRANSFER BAG 500ML W/Y CONNECTOR CE	191010001	1,040	10/10/19	OCT2022

TEST ITEMS	JUDGEMENT	REMARKS
1. TOTAL APPEARANCE	PASSED	
2. FUNCTION TEST	PASSED	
3. PYROGENITY-USP	PASSED	
4. STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

P.O. NO : 2190286

CY CHIA  
QA MANAGER

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS  
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY