



JMS SINGAPORE PTE LTD  
440 Ang Mo Kio Industrial Park 1, Singapore 569620  
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)  
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)  
Web : www.jmss.com.sg  
Reg : 197901579C

TO MEDISTA SPOL S.R.O.

Page 2 of 2

### Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO 1022231  
DATE OF ISSUE 29-DEC-20

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	201023001	3,624	23/10/20	OCT2023

TEST ITEMS	JUDGEMENT	REMARKS
1. TOTAL APPEARANCE	PASSED	
2. FUNCTION TEST	PASSED	
3. PYROGENITY-USP	PASSED	
4. STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS  
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : 2200133,2200166,2200337,2200198

CY CHIA  
QA MANAGER