



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

TO BIONIC MEDIZINTECHNIK G.M.B.H.

Page 1 of 1

Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO 1023213
DATE OF ISSUE 03-NOV-21

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-1000	CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT NP-CLP 16G CE	210910002	3,570	10/09/21	SEP2024
811-1009	CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT BSB-NP-CLP 16G CE	210903002	1,020	03/09/21	SEP2024
811-8319	BLOOD BAG CPD-SAGM (T) 450+400MLX2 5DPC IBSP-NP-CLP 16G CE	210923005	1,008	23/09/21	SEP2024
811-8319	BLOOD BAG CPD-SAGM (T) 450+400MLX2 5DPC IBSP-NP-CLP 16G CE	210924003	2,016	24/09/21	SEP2024
814-0302	TRANSFER BAG 300ML CE	210928001	10,000	28/09/21	SEP2024

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : 2210199,2210232,2210241,2210243,2210225,2210242

CY CHIA
QA MANAGER