



**JMS SINGAPORE PTE LTD**

440 Ang Mo Kio Industrial Park 1, Singapore 569620  
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)  
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)  
Web : www.jmss.com.sg  
Reg : 197901579C

**TO BIONIC MEDIZINTECHNIK G.M.B.H.**

**Certificate of Compliance**

**WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS**

**INVOICE NO** 1023920      **DATE OF ISSUE** 11-JUN-22

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-1000	CORDBLOOD COLLECTION SET CPDA-1 (S) 157ML W/INUJ PORT NP-CLP 16G CE	220510002	3,360	10/05/22	MAY2025
814-0121	TRANSFER BAG 150ML W/4 BAGS CE	220428001	2,000	28/04/22	APR2025
814-0121	TRANSFER BAG 150ML W/4 BAGS CE	220429001	1,000	29/04/22	APR2025

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

**P.O. NO :** 2220058,2220075,2220015

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS  
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

CY CHIA  
QA MANAGER



JMS SINGAPORE PTE LTD  
 440 Ang Mo Kio Industrial Park 1, Singapore 569620  
 Tel : (65) 64571176 (Main), (65) 64571144 (Sales)  
 Fax : (65) 64548500 (Main), (65) 64599564 (Sales)  
 Web : www.jmss.com.sg  
 Reg : 197901579C

PRODUCT CODE : 811-1000

LOT NUMBER : 220510002

DESCRIPTION : CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT NP-CLP 16G CE

TEST ITEMS	SPECIFICATION LIMITS	RESULTS
pH	5.0 - 6.0	PASSED 5.7
CITRIC ACID (%)	95 - 105	PASSED 99.3
SODIUM BIPHOSPHATE (%)	95 - 105	PASSED 99.8
* TOTAL CITRATE (%)	95 - 105	PASSED 101.7
DEXTROSE (%)	95 - 105	PASSED 99.8
ADENINE (%)	90 - 110	PASSED 99.3

\* TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE

THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.

DATE OF ISSUE : 11-JUN-22

CY CHIA

QA MANAGER