



**JMS SINGAPORE PTE LTD**  
440 Ang Mo Kio Industrial Park 1, Singapore 569620  
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)  
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)  
Web : www.jmss.com.sg  
Reg : 197901579C

Page 1 of 1

**TO BIONIC MEDIZINTECHNIK G.M.B.H.**

**Certificate of Compliance**

**WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS**

**INVOICE NO** 1023920  
**DATE OF ISSUE** 11-JUN-22

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-1000	CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT NP-CLP 16G CE	220510002	3,360	10/05/22	MAY2025
814-0121	TRANSFER BAG 150ML W/4 BAGS CE	220428001	2,000	28/04/22	APR2025
814-0121	TRANSFER BAG 150ML W/4 BAGS CE	220429001	1,000	29/04/22	APR2025

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS  
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

**P.O. NO :** 2220058,2220075,2220015

**CY CHIA**  
QA MANAGER



JMS SINGAPORE PTE LTD  
440 Ang Mo Kio Industrial Park 1, Singapore 569620  
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)  
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)  
Web : www.jmss.com.sg  
Reg : 197901579C

PRODUCT CODE : 811-1000

LOT NUMBER : 220510002

DESCRIPTION : CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT NP-CLP 16G CE

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH	5.0 - 6.0	5.7	PASSED
	CITRIC ACID (%)	95 - 105	99.3	PASSED
	SODIUM BIPHOSPHATE (%)	95 - 105	99.8	PASSED
	* TOTAL CITRATE (%)	95 - 105	101.7	PASSED
	DEXTROSE (%)	95 - 105	99.8	PASSED
	ADENINE (%)	90 - 110	99.3	PASSED

\* TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE

THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.

DATE OF ISSUE : 11-JUN-22

CY CHIA

QA MANAGER