



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

Page 1 of 1

TO BIONIC MEDIZINTECHNIK G.M.B.H.

Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO 1024037
DATE OF ISSUE 16-JUL-22

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-1000	CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT NP-CLP 16G CE	220510002	30	10/05/22	MAY2025
811-1000	CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT NP-CLP 16G CE	220603002	3,360	03/06/22	JUN2025
814-0107	TRANSFER BAG 1000ML CE	220606001	2,000	06/06/22	JUN2025
814-0107	TRANSFER BAG 1000ML CE	220607001	2,000	07/06/22	JUN2025
814-0302	TRANSFER BAG 300ML CE	220524001	4,000	24/05/22	MAY2025
814-0302	TRANSFER BAG 300ML CE	220527001	4,000	27/05/22	MAY2025
814-0302	TRANSFER BAG 300ML CE	220531001	2,000	31/05/22	MAY2025

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : 2220103,2220092,2220096,2220088,22200
69,2220037,2220015,2220068

CY CHIA
QA MANAGER



JMS SINGAPORE PTE LTD
 440 Ang Mo Kio Industrial Park 1, Singapore 569620
 Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
 Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
 Web : www.jmss.com.sg
 Reg : 197901579C

PRODUCT CODE : 811-1000

LOT NUMBER : 220603002

DESCRIPTION : CORD BLOOD COLLECTION SET CPDA-1 (S) 157ML W/INJ PORT NP-CLP 16G CE

TEST ITEMS	SPECIFICATION LIMITS	RESULTS
CHEMICAL CONTENTS	pH	5.6 PASSED
	CITRIC ACID (%)	95 - 105 PASSED
	SODIUM BIPHOSPHATE (%)	95 - 105 PASSED
	* TOTAL CITRATE (%)	101.8 PASSED
	DEXTROSE (%)	99.5 PASSED
	ADENINE (%)	100.0 PASSED
	5.0 - 6.0	
	90 - 110	

* TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE

THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.

DATE OF ISSUE : 16-JUL-22

CY CHIA

QA MANAGER