



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

TO MEDISTA SPOL S.R.O.

Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO 1024999
DATE OF ISSUE 08-APR-23

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
811-2577	BLOOD BAG CPDA-1 (S) 250ML IBSP-NP-CLP 16G CE	230222003	1,040	22/02/23	FEB2026
811-4012	BLOOD BAG CPDA-1 (S) 450ML CLP 16G CE	230222005	2,000	22/02/23	FEB2026
811-4032	BLOOD BAG CPDA-1 (S) 450ML IBSP-NP-CLP 16G CE	230222004	2,080	22/02/23	FEB2026
811-8442	BLOOD BAG CPD-SAGM (Q) 450+150+400MLX2 5DPC IBSP-NP-BSB-CLP 16G (DT) CE	230302002	4,032	02/03/23	MAR2026
814-0503	TRANSFER BAG 500ML CE	230206001	1,000	06/02/23	FEB2026
814-0504	TRANSFER BAG 500ML W/Y CONNECTOR CE	230207001	1,040	07/02/23	FEB2026
831-8307	BLOOD BAG T&B CPD-SAGM (T) 450+400MLX2 IBSP-NP-BSB-CLP 16G (DT) CE	230126004	2,400	26/01/23	JAN2026
831-8307	BLOOD BAG T&B CPD-SAGM (T) 450+400MLX2 IBSP-NP-BSB-CLP 16G (DT) CE	230130002	800	30/01/23	JAN2026
831-8307	BLOOD BAG T&B CPD-SAGM (T) 450+400MLX2 IBSP-NP-BSB-CLP 16G (DT) CE	230211003	840	11/02/23	FEB2026
831-8307	BLOOD BAG T&B CPD-SAGM (T) 450+400MLX2 IBSP-NP-BSB-CLP 16G (DT) CE	230307002	5,000	07/03/23	MAR2026
831-8307	BLOOD BAG T&B CPD-SAGM (T) 450+400MLX2 IBSP-NP-BSB-CLP 16G (DT) CE	230308001	5,000	08/03/23	MAR2026
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	220421002	24	21/04/22	APR2025



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

Page 2 of 2

TO MEDISTA SPOL S.R.O.

Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO DATE OF ISSUE
1024999 08-APR-23

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	230221001	3,000	21/02/23	FEB2026

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : 2220235,2220341,2220311

CY CHIA
QA MANAGER