



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

PRODUCT CODE : 831-8307

LOT NUMBER : 230430002

DESCRIPTION : BLOOD BAG T&B CPD-SAGM (T) 450+400MLX2 IBSP-NP-BSB-CLP 16G (DT) CE

CPD

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH (%)	5.0 - 6.0	5.7	PASSED
	CITRIC ACID (%)	95 - 105	99.7	PASSED
	SODIUM BIPHOSPHATE (%)	95 - 105	100.0	PASSED
	* TOTAL CITRATE (%)	95 - 105	101.0	PASSED
	DEXTROSE (%)	95 - 105	99.0	PASSED
	TOTAL SODIUM (%)	95 - 105	100.7	PASSED

* TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE

SAGM

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH (%)	4.5 - 6.5	5.3	PASSED
	SODIUM (%)	95 - 105	100.6	PASSED
	CHLORIDE (%)	95 - 105	100.3	PASSED
	DEXTROSE (%)	95 - 105	98.4	PASSED
	MANNITOL (%)	95 - 105	101.0	PASSED
	ADENINE (%)	90 - 110	101.4	PASSED

THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.

DATE OF ISSUE : 22-JUN-23

CY CHIA
QA MANAGER