



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

Certificate of Analysis

PRODUCT CODE : 811-4012

LOT NUMBER : 230522004

DESCRIPTION : BLOOD BAG CPDA-1 (S) 450ML CLP 16G CE

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH	5.0 - 6.0	5.6	PASSED
	CITRIC ACID (%)	95 - 105	100.5	PASSED
	SODIUM BIPHOSPHATE (%)	95 - 105	100.2	PASSED
	* TOTAL CITRATE (%)	95 - 105	101.8	PASSED
	DEXTROSE (%)	95 - 105	99.9	PASSED
	ADENINE (%)	90 - 110	99.4	PASSED

* TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE

THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.

DATE OF ISSUE : 22-JUN-23

CY CHIA
QA MANAGER