



JMS SINGAPORE PTE LTD  
440 Ang Mo Kio Industrial Park 1, Singapore 569620  
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)  
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)  
Web : www.jmss.com.sg  
Reg : 197901579C

Page 1 of 1

TO MEDISTA SPOL S.R.O.

### Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO      DATE OF ISSUE  
1025532          27-SEP-23

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
803-2000	CONNECTING SET 750MM - 5370061 CE	230704541	1,000	04/07/23	JUL2026
811-2577	BLOOD BAG CPDA-1 (S) 250ML IBSP-NP-CLP 16G CE	230817004	1,200	17/08/23	AUG2026
811-3038	BLOOD BAG CPDA-1 (S) 350ML IBSP-NP-CLP 16G CE	230706007	1,040	06/07/23	JUL2026
814-0503	TRANSFER BAG 500ML CE	230725001	1,000	25/07/23	JUL2026
833-8441	BLOOD BAG T&T CPD-SAGM (Q) 450+450+400+450ML IBSP-BSB-NP-CLP-WBF 16G (DT) CE	230803001	3,780	03/08/23	AUG2026
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	230811001	3,024	11/08/23	AUG2026

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS  
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : 2300094,2300149,2300120

  
CY CHIA  
QA MANAGER