



TO MEDISTA SPOL S.R.O.

Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO 1025829
DATE OF ISSUE 29-DEC-23

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
814-0504	TRANSFER BAG 500ML W/Y CONNECTOR CE	231011001	1,040	11/10/23	OCT2026
814-0611	TRANSFER BAG 600ML CE	231030001	2,000	30/10/23	OCT2026
833-8441	BLOOD BAG T&T CPD-SAGM (Q) 450+450+400+450ML IBSP-BSB-NP-CLP-WBF 16G (DT) CE	231124002	3,816	24/11/23	NOV2026
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	230811001	24	11/08/23	AUG2026
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	231113001	2,544	13/11/23	NOV2026
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	231123001	3,528	23/11/23	NOV2026

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : 2300094,2300199,2300149

CY CHIA
QA MANAGER



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

PRODUCT CODE : 833-8445

LOT NUMBER : 230811001

DESCRIPTION : BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF
16G (DT) CE

CPD

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH (%)	5.0 - 6.0	5.7	PASSED
	CITRIC ACID (%)	95 - 105	99.4	PASSED
	SODIUM BIPHOSPHATE (%)	95 - 105	99.8	PASSED
	* TOTAL CITRATE (%)	95 - 105	100.9	PASSED
	DEXTROSE (%)	95 - 105	98.8	PASSED
	TOTAL SODIUM (%)	95 - 105	100.1	PASSED

* TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE

SAGM

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH (%)	4.5 - 6.5	5.2	PASSED
	SODIUM (%)	95 - 105	100.9	PASSED
	CHLORIDE (%)	95 - 105	100.5	PASSED
	DEXTROSE (%)	95 - 105	99.5	PASSED
	MANNITOL (%)	95 - 105	99.6	PASSED
	ADENINE (%)	90 - 110	99.9	PASSED

THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.

DATE OF ISSUE : 29-DEC-23

CY CHIA
QA MANAGER