



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

TO MEDISTA SPOL S.R.O.

Page 1 of 1

Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS


INVOICE NO 1025829
DATE OF ISSUE 29-DEC-23

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
814-0504	TRANSFER BAG 500ML W/Y CONNECTOR CE	231011001	1,040	11/10/23	OCT2026
814-0611	TRANSFER BAG 600ML CE	231030001	2,000	30/10/23	OCT2026
833-8441	BLOOD BAG T&T CPD-SAGM (Q) 450+450+400+450ML IBSP-BSB-NP-CLP-WBF 16G (DT) CE	231124002	3,816	24/11/23	NOV2026
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	230811001	24	11/08/23	AUG2026
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	231113001	2,544	13/11/23	NOV2026
833-8445	BLOOD BAG T&B CPD-SAGM (Q) 450+400+450+450ML IBSP-BSB-NP-CLP-RCF 16G (DT) CE	231123001	3,528	23/11/23	NOV2026

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : 2300094,2300199,2300149


CY CHIA
QA MANAGER