



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

Page 1 of 1

TO MEDISTA SPOL S.R.O.

Certificate of Compliance

WE CERTIFY ALL LOTS MEET JMSS SPECIFICATIONS

INVOICE NO 1026124 **DATE OF ISSUE** 27-MAR-24

CODE NO	DESCRIPTION	LOT NO	QUANTITY	PROD. DATE	EXPIRY DATE
814-0135	TRANSFER BAG 150ML W/6 BAGS W/O SERIAL NUMBERS W/CLAMPS CE	240222001	1,020	22/02/24	FEB2027
814-0503	TRANSFER BAG 500ML CE	240221001	2,000	21/02/24	FEB2027
833-8441	BLOOD BAG T&T CPD-SAGM (Q) 450+450+400+450ML IBSP-BSB-NP-CLP-WBF 16G (DT) CE	231124002	36	24/11/23	NOV2026
833-8441	BLOOD BAG T&T CPD-SAGM (Q) 450+450+400+450ML IBSP-BSB-NP-CLP-WBF 16G (DT) CE	240216001	3,780	16/02/24	FEB2027

TEST ITEMS	JUDGEMENT	REMARKS
1.TOTAL APPEARANCE	PASSED	
2.FUNCTION TEST	PASSED	
3.PYROGENITY-USP	PASSED	
4.STERILITY-USP	PASSED	
5. SAFETY/TOXICITY-USP	PASSED	
6. CHEMICAL CONTENTS	PASSED	

TEST ITEMS 3-6 ARE NOT APPLICABLE TO BULK AND NON-STERILE PRODUCTS
TEST ITEM 6 IS APPLICABLE TO BLOOD BAG ONLY

P.O. NO : 2300310

CY CHIA
QA MANAGER



JMS SINGAPORE PTE LTD
440 Ang Mo Kio Industrial Park 1, Singapore 569620
Tel : (65) 64571176 (Main), (65) 64571144 (Sales)
Fax : (65) 64548500 (Main), (65) 64599564 (Sales)
Web : www.jmss.com.sg
Reg : 197901579C

PRODUCT CODE : 833-8441
LOT NUMBER : 231124002
DESCRIPTION : BLOOD BAG T&T CPD-SAGM (Q) 450+450+400+450ML IBSP-BSB-NP-CLP-WBF
16G (DT) CE

CPD

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH (%)	5.0 - 6.0	5.7	PASSED
	CITRIC ACID (%)	95 - 105	99.7	PASSED
	SODIUM BIPHOSPHATE (%)	95 - 105	99.7	PASSED
	* TOTAL CITRATE (%)	95 - 105	101.3	PASSED
	DEXTROSE (%)	95 - 105	99.1	PASSED
	TOTAL SODIUM (%)	95 - 105	100.7	PASSED

* TOTAL CITRATE = CITRIC ACID + SODIUM CITRATE

SAGM

TEST ITEMS		SPECIFICATION LIMITS	RESULTS	
CHEMICAL CONTENTS	pH (%)	4.5 - 6.5	5.1	PASSED
	SODIUM (%)	95 - 105	99.2	PASSED
	CHLORIDE (%)	95 - 105	100.2	PASSED
	DEXTROSE (%)	95 - 105	101.1	PASSED
	MANNITOL (%)	95 - 105	101.5	PASSED
	ADENINE (%)	90 - 110	99.9	PASSED

THE ABOVE RESULTS ARE WITHIN SPECIFICATIONS.

DATE OF ISSUE : 27-MAR-24


CY CHIA
QA MANAGER